

Document No.: HEK-BRP-ZL-07

Version No.: A/2

**Residual HEK293 DNA Quantitation Kit (3G)**Product No.: **1101104-1**Lot No.: **20250601**Date of Mfg.: **25 Jun. 2025**Expiration Date: **06 Mar. 2027**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$ .	PASS
Amplification efficiency of Standard curve	$E=90.0\%-110.0\%$ .	PASS
Repeatability	30 fg/ $\mu$ L HEK293 DNA Control was detected, $CV \leq 15.0\%$ .	PASS
No Template Control (NTC)	The mean value of NTC should not be more than 6.00 fg/ $\mu$ L.	PASS

Approved By Zhang LipingDAY/MONTH/YEAR 01/07/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China